

AquaTest™

Brand Name of Product	AquaTest™
Generic Name of Product	Water Test
Product Code Number(s)	AT101
Purpose of Product	To test water quality critical to proper cleaning, including pH, hardness and alkalinity.
Range of Applications for Product	Water used in the cleaning process
Key specifications of product	<ul style="list-style-type: none"> • 3- in-1 water quality test strips. <ul style="list-style-type: none"> ○ pH (2 – 12) ○ Alkalinity (0 – 500ppm CaCO3) ○ Hardness (0 – 1000 ppm CaCO3) • Result within 30 seconds. • Easy to interpret colormetric result.

Shipping & Storage

Shipping Conditions & Requirements	
Storage Conditions	
Packaging Conditions	
Shelf Life	

Instructions for Using Product

Description of Use (s)	To test water quality used in the cleaning of surgical instruments.																											
Preparation for Use																												
Diagrams (drawings, pictures):	<table border="1"> <tr> <td>pH</td> <td>2.0</td> <td>4.0</td> <td>5.0</td> <td>6.5</td> <td>8.5</td> <td>9.5</td> <td>10.5</td> <td>12.0</td> </tr> <tr> <td>Alkalinity</td> <td>0</td> <td>40</td> <td>80</td> <td>120</td> <td>180</td> <td>240</td> <td>500</td> <td></td> </tr> <tr> <td>Hardness</td> <td></td> <td>0</td> <td>50</td> <td>120</td> <td>250</td> <td>425</td> <td>1000</td> <td></td> </tr> </table>	pH	2.0	4.0	5.0	6.5	8.5	9.5	10.5	12.0	Alkalinity	0	40	80	120	180	240	500		Hardness		0	50	120	250	425	1000	
pH	2.0	4.0	5.0	6.5	8.5	9.5	10.5	12.0																				
Alkalinity	0	40	80	120	180	240	500																					
Hardness		0	50	120	250	425	1000																					
Steps for Use of Product	<ul style="list-style-type: none"> • Collect water from the cold water source into a clean container: <ol style="list-style-type: none"> 1. Dip entire strip into water for 5 seconds, then remove. 2. Shake once briskly to remove excess water from test strip. 3. Wait 20 seconds. 4. Compare each pad to the corresponding color chart to determine the value for pH, alkalinity and hardness. • Collect water from the hot water source into a clean container: <ol style="list-style-type: none"> 1. Dip entire strip into water for 5 seconds, then remove. 2. Shake once briskly to remove excess water from test strip. 3. Wait 20 seconds. 4. Compare each pad to the corresponding color chart to determine the value for pH, alkalinity and hardness. 																											
Interpretation of Results	<ol style="list-style-type: none"> 1. Color for each pad corresponds to a value which is indicated on the chart. 2. Record results on the log sheet. 3. Report any deviations from expected values. 																											
Contraindications of Test Results																												
Documentation																												
Special Warnings and Cautions																												
Disposal																												